

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130172
Name	Brent Johnson	ID#	4980 Date 5/21/09
A	Agency	AST	Phone #
Instrument Location		AST - Tok	
Alco S/N		X173017	Target Value 0.083 High Pressure 900
B			
Alco Test Values		0.083	0.084
		1 st Alco	2 nd Alco
Signature		BA [Signature] 003 12/17/09	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of Dec, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number 130172 Model DataMaster
Location TOKO AST/Tok

COS
12/17/09

Subject Information

Last Name Verif O.L. # 000000
First Name Df Cal

Operator Information

Operator Name Johnson Brent W Case 00-000
Operator # 4980 Department TOKO

Test Results

Alco Target .083 Alco Serial X173017
Nominal 0.082 Barometer 29.65 Test Type V
Dry Gas 1 0.083 Time: 12:23:14 Test Date 05/21/2009
Subject Alcohol 0.000 Time: 12:24:40 Test Time 12:21:54
Dry Gas 2 0.084 Time: 12:25:45 Report Y

Diagnostic Test Record

Serial Number 130172 Model DataMaster
Location TOKO AST/Tok

Diagnostic Test Results

Test Date 05/21/2009 Computer Okay
Test Time 12:19:23 Program Okay
Software Date 02/20/2001
Sample Chamber (C) 48
Breath Tube (C) 43
Barometer (in) 29.65
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay